



LIN. ACCEL. ROOM

DRYWALL PARTITIONS BY OTHERS
PER PROJECT REQUIREMENTS

CONCRETE STRUCTURE

OPEN LIMIT
HOME SIGNAL

DOOR CONTROL
ELECT. CAB.

STEEL TUBE

SLIDING SHIELDED DOOR

EQUIPMENT
CLOSET

SAFETY EDGE

CLOSE LIMIT
DOOR INTERLOCK

2" LEAD
2" POLYETHYLENE

STEEL TUBE

OUTSIDE DOOR
CONTROL LOCATION &
PARTITION CONFIG.
PER PROJECT REQUIREMENT

DRYWALL PARTITIONS BY OTHERS
PER PROJECT REQUIREMENTS-TYPICAL

PLAN VIEW LINAC SLIDE DOOR

**LONE STAR X-RAY
SHIELDING**
P.O. BOX 2176
SPLENDORA, TX 77372

| | |
|----------------------|-------|
| DATE | |
| SHEET | SCALE |
| 3 | NTS |
| CUSTOMER DRAWING NO. | |
| REV | |